

L.5 Development of single frame x-ray framing camera for pulsed plasma experiments

High speed gated x-ray imaging is required to obtain time resolved information in a variety of research investigations such as laser plasma interaction, capillary discharge plasma, plasma focus, tokamak devices etc. In principle, this can be accomplished by using an x-ray framing camera with a multiframe recording facility. The latter necessitates use of a stripline microchannel plate (MCP) with an array of pinholes, where a high voltage gate pulse travels along the stripline, recording each pinhole image sequentially on a phosphor screen. Such stripline MCPs are very expensive and not easily available commercially. As a simpler alternative, one can use a regular MCP, with a single pin-hole and activate the same with a short duration gating pulse to record a single frame, and obtain x ray images in consecutive time frames. A single frame x-ray framing camera with a minimum gate duration of 5ns has been developed at the Laser Plasma Division and used for measurement of x-ray emission from laser produced plasmas.




Fig. L.5.1 Single frame x-ray framing camera setup.




Fig L.5.2 A typical x-ray frame for a laser produced copper plasma and corresponding intensity profile.

The single frame x-ray framing camera consists of two parts, viz. an x ray pin-hole camera using an open ended MCP detector coupled to a CCD camera, and a variable high voltage short duration gate pulse for activating the MCP (fig. L.5.1). The camera uses a 10 μ m pin-hole aperture for

imaging the x-ray source on the MCP detector with a magnification of 6X. Design of the high voltage pulser circuit is based on self-matched transmission line [J. Upadhyay and C.P.Navathe, *Measurement Science and Technology*, 17, 25 (2006)]. The pulser circuit can generate a pulse of variable duration from 5ns to 38ns with variable amplitude from 800V to 1.2kV. In addition, the system includes a variable delay for synchronizing high voltage pulse with the event.

The performance of the system was checked by recording x-ray emission from a laser produced copper plasma. For this purpose, high power Nd:glass laser beam (30J, 2ns FWHM) was focused on a planar solid target kept at the center of a plasma chamber evacuated to $\sim 10^{-6}$ torr. A high density, high temperature plasma was produced having typical dimension of $\sim 100\mu$ m diameter. X-rays generated from this plasma were imaged onto the MCP using the x-ray pin-hole camera. The MCP output on the phosphor screen was further imaged onto a CCD camera, which in turn was controlled by a frame grabber card, triggered by the laser control unit, about 400 μ s prior to the laser pulse. [J.Upadhyay, J.A.Chakera, C.P.Navathe, P.A.Naik, A.S.Joshi, and P.D.Gupta, to appear in *Sadhana*, 2006].

A typical image of the copper plasma x-ray source produced by a 13J, 2ns laser pulse, and filtered with aluminized polycarbonate foil is shown in fig.L.5.2. The MCP gating pulse was 1.1kV / 7ns duration. A corresponding intensity profile is also shown in fig.L.5.2. From this profile, a plasma expansion size of $\sim 100\mu$ m (FWHM) was measured, which is in agreement with focal spot size of the laser used for producing the plasma source. Further, as the delay of the frame was scanned, it was observed that up to a delay of 5ns, the image continued to appear and by increasing it further, the image vanished. This shows that the x-ray emission lasts for ~ 5 ns. A reduction factor of ~ 6.5 was seen in the dark current contribution as the MCP gate pulse was decreased from 250 μ s to 5ns duration.

Contributed by:

J. Upadhyay; janky@cat.ernet.in and C.P. Navathe

L.6 Generation of hollow conic beam using an axicon mirror

Generation and characterization of hollow beams has attracted considerable attention recently because of their applications in diverse areas such as atom guiding and trapping, optical tweezers, laser machining, and generation of beams with large depth of focus. Different methods have been employed to generate hollow beams. These include use of conical lenses, holograms and transverse mode selection.